

I cannot achieve the system suitability or signal-to-noise criteria with the described chromatographic method. Can I make any adjustments?



You are here:

[EDQM FAQs](#) / [EDQM FAQs in English](#) / [EUROPEAN PHARMACOPOEIA & INTERNATIONAL HARMONISATION](#) / [General Chapters and Monographs](#) / [IMPURITIES AND CHROMATOGRAPHY](#) / I cannot achieve the system suitability or signal-to-noise criteria with the described chromatographic method. Can I make any adjustments?

Answer:

Please consult general chapter 2.2.46 ('Adjustment of chromatographic conditions' section) for a list of acceptable modifications. Please note that multiple modifications are not recommended and the system suitability criteria of the monograph and of general chapter 2.2.46 must always be met.