Does the suitability test described in general chapter 2.5.12 have to be run every time?

Answer:

The suitability test only has to be carried out once for each combination of commercial titrant/substance to be examined. If the suitability has been checked by the reagent supplier, it is not necessary to repeat the test. There is also no need to repeat the test for a new batch of substance or titrant.