

Are procedures 2.5.12 and 2.5.32 interchangeable?



You are here:

[EDQM FAQs](#) / [EDQM FAQs in English](#) / [EUROPEAN PHARMACOPOEIA & INTERNATIONAL HARMONISATION](#) / [General Chapters and Monographs](#) / [WATER - LOSS ON DRYING - SOLVENTS](#) / Are procedures 2.5.12 and 2.5.32 interchangeable?

Answer:

No, the procedure described in the monograph must be applied. If the micro-determination (2.5.32) is carried out instead of the semi-micro determination (2.5.12), or vice versa, this is considered to be the use of an alternative procedure, which would therefore require cross-validation (see the *General Notices*).