

When is a laboratory classified as ‘unsatisfactory’?



You are here

[EDQM FAQs](#) / [EDQM FAQs in English](#) / [HEALTHCARE](#) / [Blood transfusion](#) / [Blood Proficiency Testing Scheme \(B-PTS\)](#) / [Specific information related to the results and the statistical analysis of a B-PTS study](#) / [B-PTS results and statistical analysis](#) / When is a laboratory classified as ‘unsatisfactory’?

Serology and NAT studies

A laboratory is classified as ‘unsatisfactory’ if:

- one or more of the core-positive or core-negative samples is not correctly determined as ‘reactive’ or ‘non-reactive’, respectively;
- and/or the ratios interpreted from the raw data provided are not correct.

Immuno-haematology studies

A laboratory is classified as ‘unsatisfactory’ if at least one the core-positive samples is not correctly characterised/found negative.