

When is a laboratory classified as 'non-evaluable'?



You are here

[EDQM FAQs](#) / [EDQM FAQs in English](#) / [HEALTHCARE](#) / [Blood transfusion](#) / [Blood Proficiency Testing Scheme \(B-PTS\)](#) / [Specific information related to the results and the statistical analysis of a B-PTS study](#) / [B-PTS results and statistical analysis](#) / When is a laboratory classified as 'non-evaluable'?

A laboratory is classified as '**non-evaluable**' if:

- at least one result is missing;
- more than one result is inconclusive. Please note that 'inconclusive result' should be used in case of a technical failure (e.g. sampling failure, clotting, instrument failure), whereas 'doubtful' should be used for doubtful/grey zone results as implemented by the laboratory.