

PEAK SYMMETRY



'Unless otherwise stated, in a test or assay, the symmetry factor (tailing factor) of the peak used for quantitation is 0.8-1.8.'

- Does the peak symmetry requirement apply to all chromatographic procedures?
- What does 'Unless otherwise stated' mean in this context?