

Does the peak symmetry requirement apply to all chromatographic procedures?



You are here:

[EDQM FAQs](#) / [EDQM FAQs in English](#) / [EUROPEAN PHARMACOPOEIA & INTERNATIONAL HARMONISATION](#) / [General Chapters and Monographs](#) / [GENERAL CHAPTER 2.2.46 \(11.0\)](#) / [PEAK SYMMETRY](#) / Does the peak symmetry requirement apply to all chromatographic procedures?

Answer:

This requirement applies to the peak(s) used for the quantitation by external standardisation (quantitation or limit test) in any LC or GC purity test or assay. Consequently, it does not apply to analytical procedures using area normalisation as the quantitation procedure.